

## Certificate of Analysis

### Product Information:

<b>Product Code</b>	EC4421000BT
<b>Description</b>	1000 PPM @ 25.0 Deg C 442 Conductivity Solution
<b>Batch Number</b>	444/01
<b>Manufacturing Date</b>	30-Oct-2024
<b>Expiration Date</b>	30-Oct-2027

### Actual Analysis

<b>Measurement</b>	<b>Acceptance Range</b>	<b>Results</b>
Conductivity (@25degC)	1450 uS/cm +/- 1%	Pass

### Method of Standardization

This solution is standardized using conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurements.

Approved By

**ThermoFisher Scientific**  
**Water and Lab Products**  
**Quality Assurance**

This is a computer-generated document. No signature is required

Rev A