

Certificate of Analysis

Product Information:

| | |
|---------------------------|----------------------|
| Product Code | ECBU10BX |
| Description | pH 10 Buffer Tablets |
| Source Lot | W0FA0228 |
| Manufacturing Date | 28-Feb-2023 |
| Expiration Date | 28-Feb-2028 |

Actual Analysis

| | | |
|--------------------|-------------------------|----------------|
| Measurement | Acceptance Range | Results |
| pH value | 10.0 (+/- 0.05) pH | Pass |

Method of Standardization

NA

Approved By

ThermoFisher Scientific
Water and Lab Products
Quality Assurance

This is a computer-generated document. No signature is required

Rev A