

## Certificate of Analysis

### Product Information:

|                    |                     |
|--------------------|---------------------|
| Product Code       | ECBU4BX             |
| Description        | pH 4 Buffer Tablets |
| Source Lot         | M05A                |
| Manufacturing Date | 30-Sep-2021         |
| Expiration Date    | 30-Sep-2026         |

### Actual Analysis

|             |                   |         |
|-------------|-------------------|---------|
| Measurement | Acceptance Range  | Results |
| pH value    | 4.0 (+/- 0.05) pH | Pass    |

### Method of Standardization

NA

Approved By

**ThermoFisher Scientific**  
**Water and Lab Products**  
**Quality Assurance**

This is a computer-generated document. No signature is required

Rev A