

Certificate of Analysis

Product Information:

Product Code	ECBU4BX
Description	pH 4 Buffer Tablets
Source Lot	N04A
Manufacturing Date	30-Nov-2021
Expiration Date	30-Nov-2026

Actual Analysis

Measurement	Acceptance Range	Results
pH value	4.0 (+/- 0.05) pH	Pass

Method of Standardization

NA

Approved By

ThermoFisher Scientific
Water and Lab Products
Quality Assurance

This is a computer-generated document. No signature is required

Rev A