
Certificate Of Analysis

Product Information

Part Number: ECCON1413BS
Description: EUTECH 1413 μ S/CM COND POUCHES 20 PACK
Lot Number: CQ2A
Expiration Date: 10/31/2026
Method Of Analysis: CONDUCTIVITY

Actual Analysis

Acceptance Range	Results	Unit Of Measure
1413 US/CM +/- 7 US/CM AT 25 C	1413	US/CM AT 25 C

Traceability

N.I.S.T Chemical	SRM#
POTASSIUM CHLORIDE	999c

Certified By: Thermo Fisher Scientific

Water and Lab Products

Quality Assurance

Certificate Date: 12/05/24

244721-001 Rev.K

Water and Lab Products

Australia: (613) 9757-4300 In Australia: (1300) 735-295 China: (86) 21-6865-4588
Germany: (49) 6184-90-6321 India: (91) 22-6716-2261/2247 Japan: (81) 045-453-9175
North America: 1-978-232-6000 Toll Free: 1-800-225-1480 Singapore: (65) 6778-6876

Find out more at thermofisher.com/water