

Certificate of Analysis

Product Information:

Product Code	ECCON1413BT
Description	1413 uS/cm @ 25 deg C Conductivity Solution
Batch Number	035/07
Manufacturing Date	17-Jan-2025
Expiration Date	17-Jan-2028

Actual Analysis

Measurement	Acceptance Range	Results
Conductivity @ 25 deg C	1413 uS/cm +/- 1%	Pass

Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By

ThermoFisher Scientific
Water and Lab Products
Quality Assurance

This is a computer-generated document. No signature is required

Rev A