

## Certificate of Analysis

### Product Information:

<b>Product Code</b>	ECCON2764BT
<b>Description</b>	2764 uS/cm @ 25 deg C Conductivity Solution
<b>Batch Number</b>	134/02
<b>Manufacturing Date</b>	29-Mar-2024
<b>Expiration Date</b>	29-Mar-2027

### Actual Analysis

<b>Measurement</b>	<b>Acceptance Range</b>	<b>Results</b>
Conductivity (@25 deg C)	2764 uS/cm +/- 1%	Pass

### Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by a NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By

**ThermoFisher Scientific**  
**Water and Lab Products**  
**Quality Assurance**

This is a computer-generated document. No signature is required

Rev A