

Certificate of Analysis

Product Information:

Product Code	ECCON500BT
Description	500 uS/cm @ 25 deg C Conductivity Solution
Batch Number	144/04
Manufacturing Date	05-Apr-2024
Expiration Date	05-Apr-2027

Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@25 deg C)	500 uS/cm +/- 2%	Pass

Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by a NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By

ThermoFisher Scientific
Water and Lab Products
Quality Assurance

This is a computer-generated document. No signature is required

Rev A