

Certificate of Conformity

Thermo Scientific™ Finntip™ Sterile

Product Code:	94052410
Product Name:	Finntip Filter 1000 µl, sterile
Lot Number:	24254H0
Exp:	2029-09

Thermo Fisher Scientific guarantees that Thermo Scientific™ Finntip™ pipette tips meet the industry's highest standards for molding precision and clarity. Our operation system is certified and conforms to ISO 9001 and ISO 14001. Finntip pipette tips are manufactured under controlled environment and are certified by validation of the manufacturing process to be free of RNase, DNase, DNA, ATP, PCR inhibitors and endotoxin contamination.

DNA	< 30 pg
RNase	< 10 ⁻⁹ Kunitz units/µl
DNase	< 10 ⁻⁷ Kunitz units/µl
Endotoxins	< 0.06 EU/ml
ATP	< 10 ⁻¹³ mg/µl
PCR inhibitors	Not present

The sterility (SAL10⁻⁶) is ensured by irradiating this product according to ISO 11137 guidelines. Sterile products can be identified from the indicator dot on the package label. The indicator dot changes color from yellow to red under the influence of radiation.

In Joensuu, 30th September 2024



Juha Kananen
Quality Manager
Thermo Fisher Scientific Oy