

## Certificate of Conformity Finntip™ Sterile

Product Code: 94060513  
Product Name: Finntip Flex 300, sterile  
Lot Number: 21124A0  
Exp: 2026-05

We hereby confirm that these high quality Finntip™ pipette tips are manufactured under controlled environment and are certified by validation of the manufacturing process to be free of the following contaminants:

Human DNA	$< 1.9 \times 10^{-11}$ g/tip
DNase	$< 9.4 \times 10^{-4}$ U/tip
RNase	$< 6.5 \times 10^{-8}$ U/tip
Endotoxin	$< 0.05$ EU/ml

The sterility (SAL 10<sup>-6</sup>) is ensured by irradiating this product according to ISO 11137 guidelines.

In Joensuu, 28th February 2022



---

Heidi Salokangas, Quality Manager  
Laboratory Plastics Essentials  
Thermo Fisher Scientific Oy