

Certificate of Conformity Finntip™ Sterile

Product Code: 94060513
Product Name: Finntip Flex 300, sterile
Lot Number: 22169A0
Exp: 2027-06

We hereby confirm that these high quality Finntip™ pipette tips are manufactured under controlled environment and are certified by validation of the manufacturing process to be free of the following contaminants:

Human DNA	$< 1.9 \times 10^{-11}$ g/tip
DNase	$< 9.4 \times 10^{-4}$ U/tip
RNase	$< 6.5 \times 10^{-8}$ U/tip
Endotoxin	< 0.05 EU/ml

The sterility (SAL 10⁻⁶) is ensured by irradiating this product according to ISO 11137 guidelines.

In Joensuu, 5th July 2022



Heidi Salokangas, Quality Manager
Laboratory Plastics Essentials
Thermo Fisher Scientific Oy