

Certificate of Quality  
Thermo Scientific ClipTip™

Product Code: 94410610  
Product Name: ClipTip 300 Ext, Low Retention  
Lot Number: 19058B0  
Expiry date: 2024-02

Thermo Fisher Scientific guarantees that Thermo Scientific ClipTip™ Pipette tips meet the industry's highest standards for molding precision and clarity. Our operation system is certified and conforms to ISO 9001, ISO 13485 and ISO 14001. ClipTip pipette tips are manufactured under controlled environment and are certified by validation of the manufacturing process to be free of RNase, DNase, ATP and endotoxin contamination.

DNA	< 30 pg
RNase	< 10 <sup>-9</sup> Kunitz units/μl
DNase	< 10 <sup>-7</sup> Kunitz units/μl
Endotoxins	< 0.06 EU/ml
ATP	< 10 <sup>-13</sup> mg/μl

In Joensuu, 16th

June 2021



---

Heidi Salokangas  
Quality Manager  
Thermo Fisher Scientific Oy