

## Certificate of Conformity

### Thermo Scientific™ ClipTip™ Sterile


Product Code:	94420043
Product Name:	Clip Tip 12,5 Filter, sterile
Lot Number:	24093B0
Exp:	2029-04

Thermo Fisher Scientific guarantees that Thermo Scientific™ ClipTip™ pipette tips meet the industry's highest standards for molding precision and clarity. Our operation system is certified and conforms to ISO 9001 and ISO 14001. ClipTip pipette tips are manufactured under controlled environment and are certified by validation of the manufacturing process to be free of RNase, DNase, DNA, ATP, PCR inhibitors and endotoxin contamination.

DNA	< 30 pg
RNase	< 10 <sup>-9</sup> Kunitz units/μl
DNase	< 10 <sup>-7</sup> Kunitz units/μl
Endotoxins	< 0.06 EU/ml
ATP	< 10 <sup>-13</sup> mg/μl
PCR inhibitors	Not present

The sterility (SAL10<sup>-6</sup>) is ensured by irradiating this product according to ISO 11137 guidelines. Sterile products can be identified from the indicator dot on the package label. The indicator dot changes color from yellow to red under the influence of radiation.

In Joensuu,                      24th                      April                      2024

For   
 \_\_\_\_\_  
 Nathan Jenkins  
 Quality Manager  
 Thermo Fisher Scientific Oy