

Certificate of Quality Thermo Scientific™ ClipTip™

Product Code: 94420718
Product Name: ClipTip 1000, Filter Reload, Sterile, Low Retention
Lot Number: 22110B0
Exp: 2027-04

Thermo Fisher Scientific guarantees that Thermo Scientific™ ClipTip™ pipette tips meet the industry's highest standards for molding precision and clarity. Our operation system is certified and conforms to ISO 9001 and ISO 14001. ClipTip pipette tips are manufactured under controlled environment and are certified by validation of the manufacturing process to be free of RNase, DNase, ATP and endotoxin contamination.

DNA	< 30 pg
RNase	< 10^{-9} Kunitz units/ μ l
DNase	< 10^{-7} Kunitz units/ μ l
Endotoxins	< 0.06 EU/ml
ATP	< 10^{-13} mg/ μ l

The sterility ($SAL10^{-6}$) is ensured by irradiating this product according to ISO 11137 guidelines. Sterile tips can be identified from the indicator dot on the package label, product unit and Certificate of Quality. The indicator dot changes color from yellow to red under the influence of radiation.

In Joensuu, 27th April 2022



Heidi Salokangas
Quality Manager
Thermo Fisher Scientific Oy