### **thermo**scientific

# **Certificate of Analysis**

### POROS 50 HQ 1,000ML BULK PK

Product No. **1-2559-07**Lot No. **HQ150-869** 

Date of Manufacture 22JUN2021

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Labels Correct and Intact		Pass
Caps Sealed		Pass
PRODUCT APPEARANCE		
White to Off-White Media Slurry		Pass
DYNAMIC BINDING CAPACITY SOP 100040129		
5% Breakthrough with Bovine Serum Albumin	63.5 - 91.0 mg/mL	78.36 mg/mL
PROTEIN STANDARD SEPARATION SOP 100040128		
Ovalbumin Conductivity	20.40 - 21.70 mS/cm	20.7064 mS/cm
Soybean Trypsin Inhibitor Conductivity	28.00 - 30.00 mS/cm	28.6766 mS/cm
FTIR ANALYSIS SOP 3019		

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: 2 Preston Court Bedford, MA, USA

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com

For inquiries, contact us at  $\underline{cofarequests@thermofisher.com}$ 

Kim Kheav (Jul 14, 2021 11:14 EDT)

Quality Assurance
Issued 14JUL2021

## **Certificate of Analysis**

## **thermo**scientific

<2.46 CFU/mL

### POROS 50 HQ 1,000ML BULK PK

Product No. **1-2559-07** 

Lot No. **HQ150-869** 

Date of Manufacture 22JUN2021

Total Bioburden

TEST	SPECIFICATION	RESULT
Compares to reference spectrum		Pass
BIOBURDEN TEST SOP 4371392		

<= 100 CFU/mL

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: 2 Preston Court Bedford, MA, USA

Life Technologies 5781 Van Allen Way Carlsbad, CA, USA 92008 www.thermofisher.com

For inquiries, contact us at cofarequests@thermofisher.com

Kim Kheav (Jul 14, 2021 11:14 EDT)

Quality Assurance
Issued 14JUL2021