

## POROS 50 HQ 1,000ML BULK PK

Product No. 1-2559-07  
Lot No. HQ150B1180  
Date of Manufacture 12APR2024

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Labels Correct and Intact		Pass
Caps Sealed		Pass
PRODUCT APPEARANCE		
White to Off-White Media Slurry		Pass
DYNAMIC BINDING CAPACITY		
SOP 100040129		
5% Breakthrough with Bovine Serum Albumin	63.5 - 91.0 mg/mL	77.51 mg/mL
PROTEIN STANDARD SEPARATION		
SOP 100040128		
Ovalbumin Conductivity	20.40 - 21.70 mS/cm	20.8076 mS/cm
Soybean Trypsin Inhibitor Conductivity	28.00 - 30.00 mS/cm	29.0184 mS/cm
FTIR ANALYSIS		
SOP 3019		

**Pharmaceutical Grade Reagent.**  
**For Manufacturing and Laboratory Use only.**

**Manufactured in compliance with our ISO 13485 certified quality management system.**  
**Site: 2 Preston Court Bedford, MA USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*Amy Fantasia*  
**Quality Assurance**  
**Issued 14MAY2024**

## POROS 50 HQ 1,000ML BULK PK

Product No. 1-2559-07  
Lot No. HQ150B1180  
Date of Manufacture 12APR2024

TEST	SPECIFICATION	RESULT
Compares to reference spectrum		Pass
<b>BIOBURDEN TEST</b> SOP 4371392		
Total Bioburden	<= 100 CFU/mL	<10 CFU/mL

**Pharmaceutical Grade Reagent.**  
**For Manufacturing and Laboratory Use only.**

**Manufactured in compliance with our ISO 13485 certified quality management system.**  
**Site: 2 Preston Court Bedford, MA USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

Amy Fantasia  
Quality Assurance  
Issued 14MAY2024