

POROS 50 HQ 1,000ML BULK PK

Product No.1-2559-07

Lot No.HQ150B1195

Date of Manufacture15MAY2024

| TEST | SPECIFICATION | RESULT |
|---|---------------------|---------------|
| PHYSICAL INSPECTION | | |
| Labels Correct and Intact | | Pass |
| Caps Sealed | | Pass |
| PRODUCT APPEARANCE | | |
| White to Off-White Media Slurry | | Pass |
| DYNAMIC BINDING CAPACITY | | |
| SOP 100040129 | | |
| 5% Breakthrough with Bovine Serum Albumin | 63.5 - 91.0 mg/mL | 75.22 mg/mL |
| PROTEIN STANDARD SEPARATION | | |
| SOP 100040128 | | |
| Ovalbumin Conductivity | 20.40 - 21.70 mS/cm | 20.5486 mS/cm |
| Soybean Trypsin Inhibitor Conductivity | 28.00 - 30.00 mS/cm | 28.3802 mS/cm |

Pharmaceutical Grade Reagent.
For Manufacturing and Laboratory Use only.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: 2 Preston Court Bedford, MA USA

POROS 50 HQ 1,000ML BULK PK

Product No.

1-2559-07

Lot No.

HQ150B1195

Date of Manufacture

15MAY2024

| TEST | SPECIFICATION | RESULT |
|--------------------------------------|---------------|------------|
| FTIR ANALYSIS SOP 3019 | | |
| Compares to reference spectrum | | Pass |
| BIOBURDEN TEST SOP 4371392 | | |
| Total Bioburden | <= 100 CFU/mL | <10 CFU/mL |

Pharmaceutical Grade Reagent.
For Manufacturing and Laboratory Use only.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: 2 Preston Court Bedford, MA USA