



Certificate of Analysis

ZOOM[®] Focusing Buffer pH 3–7

Product No. ZB10037

Lot No. 1179564

pH Solution:

Specifications: (5.40 – 5.68)

Results: Meets Specification

Conductivity Solution:

Specifications: (1.6 – 2.6 mS/cm)

Results: Meets Specification

Refractive Index of Solution:

Specifications: (7.6 – 8.4)% Brix

Results: Meets Specification

Overall Results:

PASS

All products are for research use only. Caution: Not intended for human or animal diagnostic or therapeutic uses. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

Rick Fletcher
Director, Quality
Issued on 02 Aug 2012