

# Certificate of Analysis



## CSPD Ready-to-Use 0.25mM

Product No. T2143  
Lot No. 1212052  
Date of Manufacture 23MAY2012

TEST	SPECIFICATION	RESULT
<b>BACKGROUND</b>		
Background Luminescence (RLU)	0.200 - 1.000 RLU	0.405 RLU
<b>HPLC PURITY</b>		
Purity by HPLC 1:3	> = 80.00 %	100.0 %
Purity by HPLC 1:200	> = 95.00 %	100.0 %
<b>DEPHOSPHORYLATION</b>		
Peak Dephosphorylation Luminescence (RLU)	2500 - 7000 RLU	3201 RLU

*For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.*

**Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Bedford, MA, USA**

*Mary Reilly*

Life Technologies  
5791 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.lifetechnologies.com](http://www.lifetechnologies.com)  
[www.appliedbiosystems.com](http://www.appliedbiosystems.com)  
For inquiries, contact us at [cofarequests@lifetech.com](mailto:cofarequests@lifetech.com)

**Quality Assurance  
Issued 31DEC12**