

# Certificate of Analysis



## ABI Prism GeneScan -350 TAMRA Size Standard

Product No. 401736  
Lot No. 1307185  
Date of Manufacture 26JUL2013

TEST	SPECIFICATION	RESULT
<b>PHYSICAL INSPECTION</b>		
Correct label and lot number.		Pass
Caps are tight with no signs of leakage.		Pass
Clear solution, free from visible foreign contamination		Pass
<b>USE TEST</b>		
Each bulk lot of Genescan-350 TAMRA is tested for peak height and pattern with the ABI Prism GeneScan System		
Peak Height		Pass
Peak Pattern		Pass

*For Research Use Only. Not for use in diagnostic procedures. Expiration Date: One year from date of manufacture.*

*Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Warrington, UK*

Life Technologies  
5791 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.lifetechnologies.com](http://www.lifetechnologies.com)  
[www.appliedbiosystems.com](http://www.appliedbiosystems.com)  
For inquiries, contact us at [cofarequests@lifetech.com](mailto:cofarequests@lifetech.com)

  
Quality Assurance  
Issued 29JUL13