



Certificate of Analysis

SuperScript® One–Cycle cDNA Kit
for use with Affymetrix® One–Cycle Assays

Product No. A10752–030

Lot No. 1482095

Date of Manufacture 30Nov2013

Product Qualification

When synthesizing double stranded cDNA from 1ug Total HeLa RNA after one hour at 42°C, the product must meet the following specifications:

cDNA Synthesis Yield

The yield for 3 of 4 cDNA samples must be ≥ 500 ng.

Results: Meet specifications.

cDNA Size Range Verification By Agarose Gel Analysis

cDNA products should appear as a broad, fairly uniform smear stretching from 100 bp to >2 kb and centered at approximately 2 kb.

Results: Meet specifications

Full Length cDNA Verification by Quantitative Real–Time PCR

The ratio of the Ct values for 3' to 5' termini amplification of GAPDH should be 1.00 ± 0.10 .

Results: Meet specifications

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

Rick Fletcher
Director, Quality
Issued on 16 Nov 2013