

**Catalog Number** T10282  
**Product Name** Trypan Blue stain 0.4% \*for use with Countess® automated cell counter\*  
**CAS Number / Name** 72-57-1  
**Molecular Formula** N/A  
**Appearance** blue solution  
**Lot Number** 1896852

	LOT DATA	SPECIFICATION
<b>ASSAY</b> Osmolality	meets specification	≥290 to ≤340 mOsm/kg
<b>ASSAY</b> Sterility Testing	meets specification	negative
<b>MICROSCOPY</b> Viable Stain	meets specification	acceptable



Rachel Smith, Quality Assurance Manager

16-Nov-2016

*Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.*