

Certificate of Analysis



Dual Light Buffer A

Product No. **T2176**
Lot No. **2934C456**
Date of Manufacture **01SEP2015**

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	1.00 - 5.74 RLU	3.43 RLU
HIGH ENZYME		
High Enzyme Dilution Test- Signal/Noise	870 - 6848	2730.28

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Bedford, MA, USA

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

**Quality Assurance
Issued 24SEP15**