

Neon™ Transfection System 10 µL Kit

Product No. MPK1025
Lot No. 2K10471
Date of Manufacture 17Aug2015

Tip, naked eye inspection (Bending & length)

Specification: Comparison with standard quality product

Result: Meets specification.

Tip, Plunger working inspection

Specification: Working well

Result: Meets specification.

E Buffer, pH

Specification: $7.4 \pm 7\%$

Result: Meets specification.

R Buffer, pH

Specification: $7.4 \pm 7\%$

Result: Meets specification.

E Buffer, Conductivity

Specification: $5.70\text{ms} \pm 5\%$

Result: Meets specification.

R Buffer, Conductivity

Specification: $5.70\text{ms} \pm 5\%$

Result: Meets specification.

T Buffer, Conductivity

Specification: $4.53\text{ms} \pm 5\%$

Result: Meets specification.

Buffer Volume (Buffer, Naked eye Inspection)

E Buffer

Specification: 75 ml

Result: Meets specification.

R Buffer

Specification: 1 ml

Result: Meets specification.

T Buffer

Specification: 1 ml

Result: Meets specification.

Buffer Color

Specification: colorless and clear

Result: Meets specification.

Check the material in a Bottle bottom

Specification: PP (polypropylene)

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@lifetech.com



Nichole Reaksecker
Quality Manager
Issued on 01 Sep 2015