

## FG, Lysis Soln, 250mL Nucleic Acid Purification

Product No. **4305895**  
Lot No. **2308175**  
Date of Manufacture **31AUG2023**  
Expiration Date **20AUG2025**

TEST	SPECIFICATION	RESULT
<b>PHYSICAL INSPECTION</b>		
Physical Inspection		Pass
<b>PH</b>		
pH	5.42 - 6.17	5.86
<b>CONDUCTIVITY</b>		
Conductivity	19.0 - 27.0 mS/cm	26.79 mS/cm
<b>REFRACTIVE INDEX</b>		
Refractive Index	1.43 - 1.43	1.43

*For Research Use Only. Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system.  
Site: Warrington, UK**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*J. Heald-Green*

**Quality Assurance  
Issued 06SEP2023**