

Certificate of Analysis

3100 Capillary Array 50 cm

Product No.

4315930

Lot No.

57E076

Date of Manufacture

12- Jun-2017

	Date of Maridiacture	12-Juii-201 <i>1</i>
TEST		RESULT
CAPILLARY LEAK TEST		
Each array is tested to assure absence of leaks.		
Capillary Leak Test		PASS
CAPILLARY PLUG TEST		
Each array is tested to assure absence of plugged capillaries.		
Capillary Plug Test		PASS
, , ,		
FLUORESCENCE BACKGROUND TEST		
Each array is tested to assure background fluorescence is within specification.		
Fluorescence Background Test		PASS
U		
CAPILLARY MAPPING TEST		
Each array is tested to verify that capillary inlet and outlet spatial positioning correspon	d.	
Capillary Mapping Test		PASS

For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.



Foster City, CA

Uges Tubuy

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@lifetech.com

4366125 B