



Certificate of Analysis

7900HT Sequence Detection Systems 384-Well Spectral Calibration Kit

Product No. **4323977**

Lot No. **2303249**

Date of Manufacture **04APR2023**

Expiration Date **04APR2024**

TEST	SPECIFICATION	RESULT
------	---------------	--------

Material Test

Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.

SYBR® GREEN Dye	522 ± 5 nm	520 nm	Pass
ROX™ Dye	601 ± 5 nm	603 nm	Pass
TAMRA™ Dye	578 ± 4 nm	578 nm	Pass
NED™ Dye	570 ± 4 nm	571 nm	Pass
JOE™ Dye	548 ± 4 nm	549 nm	Pass
VIC® Dye	548 ± 4 nm	548 nm	Pass
FAM™ Dye	518 ± 4 nm	518 nm	Pass
TET™ Dye	538 ± 4 nm	537 nm	Pass

For Research Use Only. Not for use in diagnostic procedures.

**ISO13485
REGISTERED**

Singapore

Neo Xin Yi

18 APR 2023

Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@appliedbiosystems.com

Doc p/n: 100030344 Rev C