

Certificate of Analysis

7900HT Sequence Detection Systems 384-Well Spectral Calibration Kit

Product No. **4323977**
 Lot No. **2401251**
 Date of Manufacture **08JAN2024**
 Expiration Date **08JAN2025**

TEST	SPECIFICATION	RESULT
------	---------------	--------

Material Test

Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.

SYBR® GREEN Dye	522 ± 5 nm	522 nm	Pass
ROX™ Dye	601 ± 5 nm	603 nm	Pass
TAMRA™ Dye	578 ± 4 nm	577 nm	Pass
NED™ Dye	570 ± 4 nm	570 nm	Pass
JOE™ Dye	548 ± 4 nm	548 nm	Pass
VIC® Dye	548 ± 4 nm	549 nm	Pass
FAM™ Dye	518 ± 4 nm	518 nm	Pass
TET™ Dye	538 ± 4 nm	537 nm	Pass

For Research Use Only. Not for use in diagnostic procedures.



Singapore

NG MIAO LING

18 JAN 2024
 Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
 Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@appliedbiosystems.com