

Certificate of Analysis



TAQMAN ENDOGENOUS CONTROLS FG,HUMAN RPLPO

Product No. 4333761F
Lot No. 2305127
Date of Manufacture 02JUN2023

TEST	SPECIFICATION	RESULT
ANALYTICAL TESTS		
Mass Spectrometry –target Calculated Mass	+/- 0.3%	Pass
Minimum Yield Specification		Pass
Functional Test		
Real Time PCR functional test	Test Material is within +/- 1CT of reference lot	Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Warrington, UK

**Quality Assurance
Issued 09JUN2023**

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com