

Certificate of Analysis



TAQMAN ENDOGENOUS CONTROLS FG,HUMAN GUSB

Product No. 4333767F
Lot No. 2402126
Date of Manufacture 01MAR2024

TEST	SPECIFICATION	RESULT
ANALYTICAL TESTS		
Mass Spectrometry –target Calculated Mass	+/- 0.3%	Pass
Minimum Yield Specification		Pass
Functional Test		
Real Time PCR functional test	Test Material is within +/- 1CT of reference lot	Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Warrington, UK

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 04MAR2024