

TAQMAN ENDOGENOUS CONTROLS FG,HUMAN HPRT1

Product No. 4333768F
Lot No. 2406122
Date of Manufacture 26JUN2024

TEST	RESULT
ANALYTICAL TESTS	
Mass Spectrometry specification + /- 0.3%	
Mass Spectrometry (Target Calculated Mass)	Pass
Minimum Yield Specification	Pass
FUNCTIONAL TEST	
Test Material is within + /- 1CT of reference lot	
Real Time PCR functional test	Pass

*For Research Use Only.
Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Warrington, UK**

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com


**Quality Assurance
Issued 28JUN2024**