



Certificate of Analysis

TaqMan® RNase P Fast 48-Well Instrument Verification Plate

Product No. **4371439**

Lot No. **2306218**

Date of Manufacture **28SEP2023**

Expiration Date **22JUN2024**

TEST	SPECIFICATION	RESULT
The TaqMan® RNase P Fast 48-Well Instrument Verification Plate meets the Applied Biosystems 2-fold Discrimination Test and the Linearity Test specifications when tested on an Applied Biosystems StepOne™ Real-Time PCR System.		

Linearity Test	R-squared of standard curve > 0.990	Pass
2-fold Discrimination Test	(Average copy number 10K group – 3 SD) > (Average copy number 5K group + 3 SD)	Pass

For Research Use Only. Not for use in diagnostic procedures.

**ISO13485
REGISTERED**

Singapore

LINDA LIM

29 SEP 2023

Quality Assurance



Life Technologies Holdings Pte Ltd, Elock 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at
cofarequest@appliedbiosystems.com