Certificate of Analysis



Anode Buffer Container 3500 Dx Series

REF

4393925

LOT

2101137

2021-01-18

2021-09-11

TEST

SPECIFICATION

RESULT

PHYSICAL INSPECTION

Material is clear, colorless and free of particulates

Appearance - Conformance

Pass

PH TEST

Measured at 25.0 °C ± 0.3 °C

pH Value

7.6 - 8.2

Pass

CONDUCTIVITY TEST

Measured at 25.0 ℃ ± 0.1 ℃

Conductivity

1500 - 2000 μS/cm

Pass

Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.

Made in USA

Life Technologies Holdings Pte Ltd Block 33, Marsiling Industrial Estate Road 3 #07-06 Singapore 739256

IVD

Issued 19JAN2021

www.thermofisher.com

For inquiries, contact us at cofarequests@thermofisher.com