

Certificate of Analysis

ThermoFisher
SCIENTIFIC

Anode Buffer Container 3500 Dx Series

REF 4393925
LOT 2102138
 2021-02-23
 2021-10-06

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
PH TEST Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	Pass
CONDUCTIVITY TEST Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	Pass

Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.

Made in USA

 Life Technologies Holdings Pte Ltd
Block 33, Marsiling Industrial Estate Road 3 #07-06
Singapore 739256
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com




Quality Assurance
Issued 24FEB2021