Certificate of Analysis



Anode Buffer Container 3500 Dx Series

REF 4393925 LOT 2103139

√∏ 2021-04-02

2021-04-02

PHYSICAL INSPECTION

Material is clear, colorless and free of particulates

Appearance - Conformance

Pass

PH TEST

Measured at 25.0 ℃ ± 0.3 ℃

pH Value

7.6 - 8.2

Pass

CONDUCTIVITY TEST

Measured at 25.0 ℃ ± 0.1 ℃

Conductivity 1500 - 2000 μS/cm Pass

Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.

Made in USA

Life Technologies Holdings Pte Ltd
Block 33, Marsiling Industrial Estate Road 3 #07-06
Singapore 739256

€ [

IVD

Quality Assurance Issued 05APR2021

www.thermofisher.com

For inquiries, contact us at cofarequests@thermofisher.com