Certificate of Analysis



Anode Buffer Container 3500 Dx Series

REF 4393925

LOT 2104140

2021-04-28

2022-01-01

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
PH TEST		
Measured at 25.0 ℃ ± 0.3 ℃		
pH Value	7.6 - 8.2	Pass
CONDUCTIVITY TEST		
Measured at 25.0 ℃ ± 0.1 ℃		
Conductivity	1500 - 2000 μS/cm	Pass

Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.

Made in USA

Life Technologies Holdings Pte Ltd Block 33, Marsiling Industrial Estate Road 3 #07-06 Singapore 739256 www.thermofisher.com

IVD

Quality Assurance Issued 29APR2021