

# Certificate of Analysis

**ThermoFisher**  
SCIENTIFIC


## Hi-Di Formamide - 5ml

Product No. 4401457  
Lot No. 2103135  
Date of Manufacture 26APR2021  
Expiration Date 12APR2022

TEST	SPECIFICATION	RESULT
<b>CONDUCTIVITY</b>		
Measured at 23.0C +/- 0.1C		
Conductivity	19 - 34 $\mu$ S/cm	28.36 $\mu$ S/cm

*For Research Use Only. Not for use in diagnostic procedures.*

Manufactured in compliance with our ISO 13485 certified quality management system. Site:  
Bedford, MA, USA

 Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
Quality Assurance  
Issued 27APR2021