

Certificate of Analysis

ThermoFisher
SCIENTIFIC


Hi-Di Formamide - 5ml

Product No. 4401457
Lot No. 2104136
Date of Manufacture 22MAY2021
Expiration Date 11MAY2022

| TEST | SPECIFICATION | RESULT |
|----------------------------|--------------------|------------------|
| CONDUCTIVITY | | |
| Measured at 23.0C +/- 0.1C | | |
| Conductivity | 19 - 34 μ S/cm | 21.06 μ S/cm |

For Research Use Only. Not for use in diagnostic procedures.

*Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA*

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Quality Assurance
Issued 24MAY2021