



Certificate of Analysis

ThermoFisher
SCIENTIFIC

Hi-Di Formamide (5mL) 3500Dx Series

REF 4404307
LOT 2011087
 2020-12-06
 2021-12-01

TEST	SPECIFICATION	RESULT
CONDUCTIVITY TEST		
Measured at 25.0°C ± 0.1°C		
Conductivity	19 - 34 µS/cm	Pass

Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.

Made in USA

 Life Technologies Holdings Pte Ltd
Block 33, Marsiling Industrial Estate Road 3 #07-06
Singapore 739256
www.thermofisher.com

For inquiries, contact us at cofarequests@thermofisher.com




Quality Assurance
Issued 15DEC2020