



Certificate of Analysis

ThermoFisher
SCIENTIFIC

Hi-Di Formamide (5mL) 3500Dx Series

REF 4404307
LOT 2012088
 2021-02-10
 2022-01-27

TEST	SPECIFICATION	RESULT
CONDUCTIVITY TEST Measured at 25.0 °C ± 0.1 °C		
Conductivity	19 - 34 µS/cm	Pass

Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.

Made in USA

 Life Technologies Holdings Pte Ltd
Block 33, Marsiling Industrial Estate Road 3 #07-06
Singapore 739256
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com




Quality Assurance
Issued 11FEB2021