



# Certificate of Analysis

**ThermoFisher**  
SCIENTIFIC

## Hi-Di Formamide (5mL) 3500Dx Series

**REF** 4404307  
**LOT** 2104091  
 2021-05-19  
 2022-05-11

TEST	SPECIFICATION	RESULT
<b>CONDUCTIVITY TEST</b> Measured at 25.0 °C ± 0.1 °C		
Conductivity	19 - 34 µS/cm	Pass

*Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.*

**Made in USA**

 Life Technologies Holdings Pte Ltd  
Block 33, Marsiling Industrial Estate Road 3 #07-06  
Singapore 739256  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



  
**Quality Assurance**  
**Issued 20MAY2021**