



Certificate of Analysis

ThermoFisher
SCIENTIFIC

Cathode Buffer Container 3500 Dx Series

REF 4408258
LOT 2104148
 2021-05-07
 2022-01-22

TEST	SPECIFICATION	RESULT
VISUAL INSPECTION		
Clear, particulate free, viscous solution		Pass
PH TEST Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.9315
CONDUCTIVITY Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1790.5 µS/cm

Manufactured in compliance with our ISO 13485 certified quality management system and 21 CFR 820 quality system regulations.

Made in USA

 Life Technologies Holdings Pte Ltd
Block 33, Marsiling Industrial Estate Road 3 #07-06
Singapore 739256
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Quality Assurance
Issued 10MAY2021