

Certificate of Analysis

thermoscientific

POROS XQ Strong Anion Exchange Resin, 1,000 mL

Product No. 4467818
Lot No. XQ-074
Date of Manufacture 14NOV2020

TEST	SPECIFICATION	RESULT
PRODUCT APPEARANCE		
White to Off-White Slurry		Pass
FRONTAL CAPACITY SOP 4467917		
5% Breakthrough with BSA	≥ 140.0 mg/mL	143.51 mg/mL
IONIC CAPACITY SOP 4467915		
Ionic Capacity ($\mu\text{mol/mL}$)	80.0 - 130.0	86.25
PROTEIN SEPARATION SOP 4467916		
Holo-Transferrin Retention	11.4 - 15.2 mS/cm	13.4950 mS/cm
Ovalbumin Retention	20.0 - 25.0 mS/cm	21.9880 mS/cm
Soybean Trypsin Inhibitor Retention	27.0 - 34.0 mS/cm	29.8110 mS/cm
FTIR ANALYSIS SOP BED3019		

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site: 2
Preston Court Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Sandra Gil
Quality Assurance
Issued 10DEC2020

Certificate of Analysis

thermo**scientific**

POROS XQ Strong Anion Exchange Resin, 1,000 mL

Product No. 4467818
Lot No. XQ-074
Date of Manufacture 14NOV2020

TEST	SPECIFICATION	RESULT
Compares to Reference Spectrum		Pass
TOTAL BIOBURDEN SOP 4371392		
Total Bioburden	<= 100 CFU/mL	<2.58 CFU/mL

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site: 2
Preston Court Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Sandra Gil
Quality Assurance
Issued 10DEC2020