

Certificate of Analysis



DYNALIGHT RAPIDGLOW RTU 100ML

Product No. 4475406
Lot No. 2312066
Date of Manufacture 27JUN2023

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	≤ 100 RLU	2.4 RLU
SIGNAL TEST/SIGNAL STD		
% ACTIVITY	112 - 182 %	172.5 %

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA**

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Quality Assurance
Issued 23JAN2024