

# Certificate of Analysis



## DYNALIGHT RAPIDGLOW RTU 100ML

Product No. 4475406  
Lot No. 2403067  
Date of Manufacture 27JUN2023

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	<= 100 RLU	2.4 RLU
SIGNAL TEST/SIGNAL STD		
% ACTIVITY	112 - 182 %	172.5 %

*For Research Use Only. Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*Amy Fantasia*  
**Quality Assurance  
Issued 27MAR2024**