

# Certificate of Analysis



## DYNALIGHT RAPIDGLOW RTU 100ML

Product No. 4475406  
Lot No. 2403068  
Date of Manufacture 10JAN2024

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	<= 100 RLU	5.7 RLU
Signal Test/Signal STD		
% ACTIVITY	112 - 182 %	166.2 %

**For Research Use Only.**  
**Not for use in diagnostic procedures.**

**Manufactured in compliance with our ISO 13485 certified quality management system.**  
**Site: Bedford, MA USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
**Quality Assurance**  
**Issued 25APR2024**