

Certificate of Analysis



DYNALIGHT RAPIDGLOW RTU 100ML

Product No. 4475406
Lot No. 2407070
Date of Manufacture 26OCT2023

| TEST | SPECIFICATION | RESULT |
|-------------------------------|---------------|---------|
| BACKGROUND | | |
| Background Luminescence (RLU) | <= 100 RLU | 4.2 RLU |
| Signal Test/Signal STD | | |
| % ACTIVITY | 112 - 182 % | 147.0 % |

*For Research Use Only.
Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA, USA**

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


**Quality Assurance
Issued 13AUG2024**