



Certificate of Analysis

ZOOM® Carrier Ampholytes 3–10

Product No. ZM0021

Lot No. 548544B

Description

Clear solution of polyamino–polysulfonic–polycarboxylic acids, 40% (w/v) in water, sterile filtered.

pH

Specification: 6.9 ± 0.4

Results: Meets specification.

Conductivity

Specification: 400 to 890 $\mu\text{S}/\text{cm}$

Results: Meets specification.

pH–range

Specification: $3.3 \pm 0.5 - 10.0 \pm 0.6$

Results: Meets specification.

pH–gradient formed

Specification: Not less than 83%

Results: Meets specification.

All products are for research use only. Caution: Not intended for human or animal diagnostic or therapeutic uses. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

Rick Fletcher
Director, Quality
Issued on 14 May 2012