


Certificate of Analysis

96-Well Spectral Calibration Plate with ABY® Dye for Multiplex qPCR



A24738

2310066

07 NOV 2023

07 NOV 2024

TEST	SPECIFICATION	RESULT
Material Test		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
ABY® Dye	573 ± 5 nm	575 nm

For Research Use Only. Not for use in diagnostic procedures.

**ISO13485
REGISTERED FACILITY**

Farah Dina

22 NOV 2023

A handwritten signature in blue ink, appearing to read 'Farah Dina', is written over a horizontal line.

Quality Assurance

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256

To obtain a Certificate of Analysis on-line go to www.thermofisher.com or email us at
QA-SG_COA_Request@lifetech.com