Certificate of Analysis



Cathode Buffer Container 3500 Series for Protein Quality Analysis

Product No.

A31279

Lot No.

2005018

Date of Manufacture

08JUN2020

Expiration Date

TEST

20FEB2021

SPECIFICATION

RESULT

PHYSICAL INSPECTION

Material is clear, colorless and free of particulates

Appearance - Conformance

Pass

PH TEST

Measured at 25.0°C ± 0.3°C

pH Value

7.6 - 8.2

7.979

CONDUCTIVITY TEST

Measured at 25.0°C ± 0.1°C

Conductivity

1500 - 2000 µS/cm

1787 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA, USA

Life Technologies 5781 Van Allen Way Carlsbad, CA, USA 92008 www.thermofisher.com For inquiries, contact us at cofarequests@thermofisher.com **Quality Assurance** Issued 09JUN20